




2015 003009S

重庆市疾病预防控制中心



检测结果报告

渝疾控消字[2015]807号

第1页/共1页

样品名称: <u>医疗废物高温蒸汽处理设施监测菌片</u>	商 标: <u>——</u>
生产单位: <u>——</u>	样品来源: <u>送检</u>
被采样单位: <u>——</u>	样品数量: <u>2片</u>
采样单位: <u>——</u>	采 样 人: <u>——</u>
送样单位: <u>重庆智得热工工业有限公司</u>	送 样 人: <u>陈涛</u>
样品状态及包装: <u>白色纸片</u>	收样日期: <u>2015年9月9日</u>
生产日期/批号: <u>——</u>	检验日期: <u>2015年9月11日</u>
检验项目: <u>压力蒸汽灭菌效果</u>	
检测依据: <u>卫生部《消毒技术规范》(2002年版) 3.17.2.1</u>	

检测结果:

该单位送检的1具医疗废物高温蒸汽处理设施(Gient-MWI2500)监测菌片(嗜热脂肪杆菌芽胞 ATCC7953 $5.0 \times 10^5 \sim 5.0 \times 10^6$ cfu/片,由中国军事医学科学院消毒检测中心生产)2片,接种的溴甲酚紫葡萄糖蛋白胨水培养基经培养均无菌生长。

(以下空白)

检测者:

何玉明

复核者:

李兵

(盖章)

2015年9月

专用章

签发人:

陈涛



Chongqing Center for Disease Control and Prevention



Report of Test Results

YJKXZ[2015] No. 807

Page 1 of 1

Name of sample: Tablets for monitoring the facility for disposing medical wastes by high temperature steam

Trademark: -----

Manufactured by: -----

Sample source: delivered for test

Sampled organization: -----

Quantity of samples: 2 tabs

Sampling organization: -----

Sampled by: -----

Delivering organization: Chongqing Gient Heating Industry Co., Ltd.

Delivered by: Chen Tao

Sample status and package: White paper

Date of reception: September 9, 2015

Date of manufacture/batch No.: -----

Date of inspection: September 11, 2015

Test item: Disinfection by high-pressure steam

Test standard: Ministry of Health, Technical Standard for Disinfection (2002 Edition) 3.17.2.1

Test results:

The tablet with bacteria (spores of bacillus stearothermophilus ATCC7953 $5.0 \times 10^5 \sim 5.0 \times 10^6$ cfu/tab, produced by the Disinfection Testing Center of the Academy of Military Medical Sciences of China), delivered by the said organization for monitoring the facility (Gient-MWI2500) for disposing medical wastes by high temperature steam, has been inoculated with bromcresol purple glucose peptone water matrix and found free of bacteria growth after being cultured.

(Blank below)

Special Seal for Report of Test Results of Chongqing Center for Disease Control and Prevention (seal)



September 11, 2015

Tested by: 何玉明
(signature)

Reviewed by: 张
(signature)





重庆华电翻译有限公司

CHONGQING HUADIAN TRANSLATION CO., LTD.

Translation Certificate

This is to certify that Chongqing Gient Heating Industry Co., Ltd. entrusted our company with translation and review from Chinese to English of *Report of Test Results* (one copy) on September 11, 2015.

Chongqing Huadian Translation Co., Ltd.



September 11, 2015

地址:重庆市江北区红石路5号北部尚座A座21-19 电话: +86 (023) 61826907 / 8 / 9

传真: +86 (023) 61826906

E_mail: hdtrans@vip.163.com <http://www.hdtrans.com>



营业执照

(副本)

渝江 注册号 500105000099626 1-1-1

名称	重庆华电翻译有限公司
类型	有限责任公司
住所	重庆市江北区红石路5号北部尚座A座21-19号
法定代表人	文仕俊
注册资本	壹佰万元整
成立日期	1998年07月15日
营业期限	1998年07月15日至永久
经营范围	电力技术服务及翻译服务；商品信息咨询；境内劳务派遣；企业营销策划。*[国家法律法规禁止的不得经营，国家法律、行政法规规定应经许可方可从事的经营，在未取得许可前不得从事经营]



登记机关



2015年02月10日

提示：每年1月1日至6月30日通过企业信用信息公示系统报送上一年度年度报告并公示，不另行通知。

gsxt.cqgs.gov.cn