



广东省微生物分析检测中心  
GUANGDONG DETECTION CENTER OF MICROBIOLOGY

# 分析检测报告

REPORT FOR ANALYSIS

报告编号	2014SP0850D
Report No.	
样品名称	MWS 型医疗废物高温蒸汽处理装置
Name of Sample	Steam-based medical waste treatment equipment : Autoclave models of MWS
委托单位	中国环境保护产业协会
Applicant	China Association of Environmental Protection Industry
检测地点	广东省微生物分析检测中心
Place for Test	Guangdong Detection Center of Microbiology

地址: 广州市先烈中路 100 号大院 59 号楼  
Address: Building 59, No.100 Central Xian Lie Road, Guangzhou, China  
邮政编码: 510070  
Postcode:  
电话号码: (020)87137666  
Tel:  
传真号码: (020)87137668  
Fax:  
网址: www.gddcm.com  
Website:



# 广东省微生物分析检测中心

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

## 分析检测报告

REPORT FOR ANALYSIS

样品名称 <u>MWS 型医疗废物高温蒸汽处理装置</u> Name of Sample <u>Steam-based medical waste treatment equipment :Autoclave models of MWS</u>	接样方式及数量 <u>送检 1 台</u> Way and Quantity of Reception <u>One sample submitted</u>
样品规格及批号 <u>MWS800</u> Spec and Lot No of Sample <u>MWS800</u>	样品状态和特性 <u>试验包</u> State and Characteristic <u>Test Bag</u>
委托单位 <u>中国环境保护产业协会</u> Applicant <u>China Association of Environmental Protection Industry</u>	接样日期 <u>2014/04/03</u> Sample Received Date <u>2014/04/03</u>
检测项目 <u>灭菌效果</u> Test Item <u>Sterilization Efficacy</u>	检测日期 <u>2014/04/03 to 2014/04/05</u> Testing Period <u>2014/04/03 to 2014/04/05</u>
报告编号 <u>2014SP0850D</u> Report No. <u>2014SP0850D</u>	签发日期 <u>2014/04/10</u> Date for Reporting <u>2014/04/10</u>
检测方法 <u>GB 15981-1995</u> Test Method <u>GB 15981-1995</u>	

## 分析检测结果

### Results

试验菌株 Type of test germ: 嗜热脂肪杆菌芽孢 ( <i>Bacillus stearothermophilus</i> ) ATCC7953						
结果 Test Result  灭菌周期 Sterilization Cycle No.	试验包 1 Test Bag No.1		试验包 2 Test Bag No.2		试验包 3 Test Bag No.3	
	菌片 1 Spore tablet No.1	菌片 2 Spore tablet No.2	菌片 1 Spore tablet No.1	菌片 2 Spore tablet No.2	菌片 1 Spore tablet No.1	菌片 2 Spore tablet No.2
灭菌周期 1 Cycle No.1	紫色 Purple	紫色 Purple	紫色 Purple	紫色 Purple	紫色 Purple	紫色 Purple
灭菌周期 2 Cycle No.2	紫色 Purple	紫色 Purple	紫色 Purple	紫色 Purple	紫色 Purple	紫色 Purple
结论: 灭菌合格 Conclusion: Test result show sterilization efficacy is qualified.						
(以下空白, blank below)						
备注 Remarks	1) 紫色表示阴性,黄色表示阳性; Purple color indicating negative result,yellow color indicating positive result. 2) 灭菌过程由客户完成 Sterilization process is completed by the customer. 3) 生产厂家 Manufacturer: 重庆智得热工工业有限公司 Chongqing Gient Heating Industry Co.,Ltd.					

制表: 刘伟君  
Editor

审核: 李良秋  
Verifier

批准: 李良秋  
Approver

盖章: 检测专用章  
Official Seal





# 广东省微生物分析检测中心

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

## 分析检测报告

REPORT FOR ANALYSIS

报告编号 (Report No.): 2014SP0850D

No.68519558

### 注意事项

Notice Items

1. 检测报告无本单位检测专用章、骑缝章无效。  
Test report is invalid if not affixed with Authorized Stamp of Test and Paging Seal.
2. 检测报告无审核人、批准人签字无效。  
Test report is invalid without signature of verifier and approver.
3. 检测报告涂改增删无效。  
Test report is invalid if being supplemented, elected or altered.
4. 未经本单位书面许可不得部分复制检测报告 (全部复制除外)。  
Without prior written permission of the laboratory, the test report cannot be reproduced, except in full.
5. 除非另有说明, 本报告检测结果仅对测试样品负责。  
Unless otherwise stated, the results shown in this test report refer only to the sample(s) tested.
6. 对检测报告若有异议, 应于收到报告之日起十五天内向检测单位提出。  
Objections to the test report must be submitted to us within 15 days.
7. 样品为客户送检时, 样品来源信息由客户提供, 实验室不负责其真实性。  
If the sample is submitted by the customer for detection, sample source information is provided by the customer and the laboratory is not responsible for its authenticity.

