



Chongqing Centers for
Disease Control and Prevention(CDC)
Report of Test Results

中国认可
国际互认
TESTING
CNASL1095

YJKXZ[2015] No.795

Page 1 of 1

Name of sample: Tablets for monitoring 2 units of MWI2500 system for disposing medical wastes by high temp. & press. steam Trademark: ----

Manufactured by: ---- Sample source: delivered for test

Sampled organization: ---- Quantity of samples: 20 tabs.

Sampling organization: ---- Sampled by: ----

Delivering organization: Chongqing Gient Heating Industry Co., Ltd. Delivered by: Yang Chao

Sample status and package: white paper Date of reception: Dec. 17. 2015

Date of manufacture/batch No.: ---- Date of inspection: Dec. 17. 2015

Test item: Sterilization Efficacy for MWI2500 System

Test standard: Ministry of Health, Technical Standard for Disinfection (2002 Edition) 3.17.2.1

Test results:

The tablet with bacteria (B.stearothermophilus spores ATCC7953 $5.0 \times 10^5 \sim 5.0 \times 10^6$ cfu/tab, produced by the Beijing Sihuan Disinfection Co.Ltd), delivered by the said organization for monitoring 2 units of MWI2500 system(unit A & B) for disposing medical waste by high temperature and pressure steam. Each unit with continuing 5 cycles, 2 pieces per cycle, and total sample quantity is 20 pieces. The samples has been inoculated with bromcresol purple glucose peptone water matrix and found free of bacteria growth after being cultured or > 8 log reduction of bacterial spores.



Tested by: 同
(signature)

Reviewed by: 朱
(signature)

Issued by: 李
(signature)